© CONTEC Ver.1.09

# High-Resolution Analog I/O Terminal for USB2.0

AIO-160802GY-USB



\* Specifications, color and design of the products are subject to change without notice.

This product is a USB2.0 compatible terminal module that extends the analog I/O function of USB port of PCs. 8ch/16bits analog inputs and 2ch/16bits analog outputs are employed and signal lines can be directly connected to the screw terminals in the system. Compact design to match Note PCs and excellent in mobility as operation is powered by USB bus.

Windows/Linux device driver is supported with the product.

- \* The contents in this document are subject to change without notice.
- \* Visit the CONTEC website to check the latest details in the document.
- \* The information in the data sheets is as of July, 2024.

### **Features**

 Equipped with the analog input 8ch, analog output 2ch and digital I/O 4ch each

Equipped with the analog input (4µsec/ch, 16bit, 8ch), analog output (10µsec, 16bit, 2ch) and digital I/O (4ch each, LVTTL level).

- Compatible to USB2.0/USB1.1 and not necessary to power this product externally as the bus power is used

Compatible to USB2.0/USB1.1 and capable to achieve high speed transfer at HighSpeed (480 Mbps). Not necessary to power this product externally as the bus power of USB is used.

Equipped with the buffer memory which can be used in either FIFO or ring format

This product includes buffer memory (8K data for analog input and 1K data for output) which can be used in either FIFO or ring format. You can perform analog I/O in the background, independent of software and the current status of the PC.

- Sampling/generating can be driven by a clock or by various triggers Sampling/generating can be started and stopped by software or by an external trigger (timing controlled by an externally input control signal). The sampling/generating period can be controlled by the internal clock (high-precision timer included on the board) or by an external clock (externally input control signal).
- Terminal connector facilitating wiring Wiring is easy as the terminal connector (screw type) is used.
- Software-based adjustment function

Adjustment of analog input/output can be all performed by software. Apart from the adjustment information prepared before shipment, additional adjustment information can be stored according to the use environment.

- Windows/Linux support device driver

Using the device driver API-TOOL makes it possible to create applications of Windows/Linux. In addition, a diagnostic program by which the operations of hardware can be checked is provided.

### **Included Items**

Product...1 Interface connector...2 USB Cable(1.8m)...1 USB Cable Attachment...1 Please read the following...1

## **Specifications**

**Function specification** 

ltem	Specifications
Analog input	
Isolated specification	Unisolated
Input type	Single-Ended Input
Number of input channels	8ch
Input range	Bipolar ±10V
Maximum input voltage	±20V
Input impedance	$1M\Omega$ or more
Resolution	16bit
Non-linear error *1*2	±12LSB
Conversion speed	4µsec/ch (Max) *3
Buffer memory	8K data
Conversion start trigger	Software / external trigger
Conversion stop trigger	Number of sampling times / external trigger / software
External start signal	LVTTL level (Selecting the rising / falling edge to the DI00-pin by the software) *4
External stop signal	LVTTL level (Selecting the rising / falling edge to the DI01-pin by the software) *4
External clock signal	LVTTL level (Selecting the rising / falling edge to the DI02-pin by the software) *4
Analog output	1 , 3 5 5 7
Isolated specification	Unisolated
Number of output channels	2ch
Output range	Bipolar ±10V
Absolute max. output currency	±1mA
Output impedance	1Ω or less
Resolution	16bit
Non-Linearity error *1	±12LSB
Conversion speed	10µsec (Max.) *3
Buffer memory	1K data
Conversion start trigger	Software / external trigger
Conversion stop trigger	Number of sampling times / external trigger/software
External start signal	LVTTL level (Selecting the rising / falling edge to the DI00-pin by the software) *2
External stop signal	LVTTL level (Selecting the rising / falling edge to the DI01-pin by the software) *2
External clock signal	LVTTL level (Selecting the rising / falling edge to the DI02-pin by the software) *2
Digital I/O	1
Number of input channels	Unisolated input 4ch (LVTTL positive logic) *5*6
Number of output channels	Unisolated output 4ch (LVTTL positive logic)
USB	
Bus specification	USB Specification 2.0/1.1 standard
USB transfer rate	12Mbps (Full-speed), 480Mbps (High-speed) *7
Power supply	Bus power
Common section	'
Number of terminals used at the same time	127 terminals (Max.) *8
Power consumption	5VDC 470mA (Max.)
Physical dimensions (mm)	64(W) x 62(D) x 24(H) (exclusive of protrusions)
Weight	90g (Not including the USB cable, attachment)
Attached cable	USB cable 1.8m


AIO-160802GY-USB 1

- \*1 The non-linearity error means an error of approximately 0.1% occurs over the maximum range at 0°C and 50°C ambient temperature. The error can be reduced by calibrating under the actual temperature conditions.
- \*2 When using the signal source equipped with the high-speed operational amplifier.
- \*3 This numerical displays the conversion speed for A/D, D/A converter. The minimum executable sampling cycle depends on the operating condition of the terminal.
- \*4 The analog input/output are shared, so cannot trigger set different trigger settings for the same terminal.
- \*5 You cannot use both the DI00 / DI01 / DI02-pin of digital input feature and the external start / stop signal / external clock input simultaneously.
- \*6 Each input accept TTL (5VDC) level signals.
- \*7 The USB transfer speed depends on the host PC environment used (OS and USB host controller).
- \*8 As a USB hub is also counted as one device, you cannot just connect 127 USB terminals.

#### **Installation Environment Requirements**

Item	Specifications	
Operating ambient temperature	0 - 50°C	
Operating ambient humidity	10 - 90%RH (No condensation)	
Floating dust particles	Not to be excessive	
Corrosive gases	None	
Standard	VCCI Class A, FCC Class A, CE Marking (EMC Directive Class A, RoHS Directive), UKCA	

# **Physical Dimensions**



# **Optional Products**

Product Name	Model type	Note
14pin Screw Terminal Connector Set	CN6-Y14	6 pieces
Bracket for USB I/O Terminal products	BRK-USB-Y	

Visit the CONTEC website for the latest optional products.

# **Support Software**

Name	Contents	How to get
Windows Version Analog I/O Driver software API-AIO(WDM)	The API-AIO(WDM) is the Windows version driver software that provides products in the form of Win32 API functions (DLL). Various sample programs such as Visual Basic and Visual C++, etc and diagnostic program useful for checking operation is provided.	Download from the CONTEC website *1
Linux Version Analog I/O Driver software API-AIO(LNX)	This is the Linux version driver software provided in API function formats. The software includes various sample programs such as gcc (C, C++) and Python programs.	Download from the CONTEC website *1
Software Development Tool Kits (SDK) and Support Software	In addition to the device drivers, we offer many software programs for using CONTEC devices in an easier manner.	Download from the CONTEC website *2

 $^{*1}$  Download the files from the following URL


https://www.contec.com/download/

\*2 For supported software, search the CONTEC website for this product and view the product page. https://www.contec.co

### **External Connectio**

### Using the On-terminal Connector

To connect an external device to this terminal, plug the cable from the device into the interface connector (CN1, CN2) shown below.



Layout on the Interface Connector(CN1, CN2)

Layout on the interface connector				
	CN2			
AGND	1			
AI 07	2			
Al 06	3			
AI 05	4			
Al 04	5	]		
AGND	6	2 3 4 5 6 7 8 9 10 11 12 12 13		
AGND	7	9		
AI 03	8	8		
AI 02	9	10		
AI 01	10	12   1		
AI 00	11	13		
AGND	12			
N.C.	13			
AGND	14			

CN1			
	14	AGND	
	13	AO 00	
	12	AGND	
14	11	AO 01	
9 11 12	10	DI 00	
10	9	DI 01	
	8	DI 02	
	7	DI 03	
	6	DGND	
10 9 8 7 6 5 4 4 3 2 1	5	DO 00	
	4	DO 01	
	3	DO 02	
	2	DO 03	
	1	DGND	

AI 00 - AI 07	00 - Al 07 Analog input signal. The numbers correspond to channel numbers.	
AO 00 - AO 01	Analog output signal. The numbers correspond to channel numbers.	
AGND	Common analog ground for analog I/O signals.	
DI 00 - DI 03	Digital input signal. The numbers correspond to input bits.	
DO 00 - DO 03 Digital output signal. The numbers correspond to output bits.		
DGND Common digital ground for digital I/O signals		

# **∴** CAUTION

- Do not connect any of the outputs and power outputs to the analog or digital ground. Neither connect outputs to each other. Doing either can result in a fault.
- If analog and digital ground are shorted together, noise on the digital signals may affect the analog signals.
   Accordingly, analog and digital ground should be separated.


#### Cable connection

When connecting the product to an external device, you can use the supplied connector plug.

For wiring, strip off approximately 9 - 10mm of the covered part of a wire rod and then insert it to the opening. After the insertion, secure the wire rod with screws. Compatible wires are AWG 28 - 16.



Removing the connector plug by grasping the cable can break the wire.



AIO-160802GY-USB 2

# **Connecting Analog Input Signal**

Analog signal input types are divided into single-ended input and differential input. This product uses single-ended input fixed. The following examples show how to connect analog input signals using a flat cable and a shielded cable.

### Single-ended Input


#### Connection example with flat cable

The following figure shows an example of flat cable connection. Connect separate signal and ground wires for each analog input channel on CN2.



# Connection example with shield cable

The following figure shows an example of shield cable connection. Use shielded cable if the distance between the signal source and product is long or if you want to provide better protection from noise. For each analog input channel, connect the core wire to the signal line and connect the shielding to ground.



#### **⚠** CAUTION

- If the signal source contains over 1MHz signals, the signal may effect the cross-talk noise between channels
- If the product and the signal source receive noise or the distance between the product and the signal source is too long, data may not be input properly.
- An input analog signal should not exceed the maximum input voltage (relate to the product analog ground).
   If it exceeds the maximum voltage, the product may be damaged.
- Connect all the unused analog input channels to analog ground.
- The signal connected to an input channel may fluctuate after switching of the multiplexer. In this case, the
  cable between this product and the signal source can be shortened or a buffer with a high-speed amplifier can
  be placed between the product and the signal source in order to reduce the fluctuation.

### **Connecting Analog Output Signal**

#### Connection example with flat cable


This section shows how to connect the analog output signal by using a flat cable or a shielded cable.

The following figure shows an example of flat cable connection. Connect the signal source and ground to the CN1 analog output.



#### Connection example with shield cable

The following figure shows an example of shield cable connection. Use shielded cable if the distance between the signal source and product is long or if you want to provide better protection from noise. For each analog input channel on CN1, connect the core wire to the signal line and connect the shielding to ground.



#### **⚠** CAUTION

- If this product or the connected wire receives noise, or the distance between this product and the target is long, data may not be outputted properly.
- For analog output signal, the current capacity is ±1mA (Max). Check the specification of the connected device before connecting this product.
- Do not short the analog output signal to analog ground, digital ground, and/or power line. Doing so may damage this product.
- Do not connect an analog output signal to any other analog output, either on this product or on an external device, as this may cause a fault on this product.


When the USB cable is plugged, the analog output signal will be -10V.

# **Connecting Digital I/O Signals**


The digital I/O signal can be used .as the control signal (external trigger input signal and sampling clock input signal, etc.), too. The following sections show examples.

All the digital I/O signals are LVTTL (3.3VDC) level signals.

#### **Digital Input Connection**




### Digital Output Connection




#### **!** CAUTION

Do not short the output signals to analog ground, digital ground, and/or power line. Doing so may damage the terminal. Each input accepts 5V TTL level signals.

# Circuit Block Diagram



# **Component Name**



No.	Name	No.	Name
1	Interface Connector	3	USB Type-A Connector
2	LED Indicator		

AIO-160802GY-USB 3